

CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME





Lab Name Govt. PHC, Pasu Mathur Lab No 13042

Constituent Group Chemistry III Date of Result Entered: 19/04/2023

PT item Lyophilized human serum based Date of Report Published : 06/05/2023

SI.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your	SDI	- 11	
SI.NO						CV	SD	Value	וחפ	U	
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1832	119.45	6.04	7.21	113 mg/dL	-0.89	0.34	
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1820	2.16	8.09	0.18	1.9 mg/dL	-1.49	0.01	
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1750	111.61	5.19	5.79	124 mg/dL	2.14	0.28	

SDI Range	Interpretation				
Within -1.00 to +1.00	Excellent.				
Within ±1.01 to ±2.00	Good.				
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.				
Beyond ±3.0	Unacceptable performance. Action Signal.				

LAB ADDRESS:

Govt. PHC, Pasu Mathur PASU MATHUR Pasu Mathur TAMILNADU635083

> Coordinator Contact Details: Email:clinqc@cmcvellore.ac.in Contact Number: 0416-2283102

Dr. Pamela Christudoss CMC EQAS Coordinator Christian Medical College, Vellore

Panela Christudoss

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

******** End of Report *******