

CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME





Lab Name Govt. PHC, Pasu Mathur Lab No 13042

Constituent Group Chemistry III Date of Result Entered : 20/03/2023

PT item Lyophilized human serum based Date of Report Published : 04/04/2023

SI.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your	SDI	- 11	
SI.NO						CV	SD	Value	SDI	U	
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1878	91.48	7.46	6.82	89.86 mg/dL	-0.24	0.31	
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1851	1.33	9.08	0.12	1.2 mg/dL	-1.07	0.01	
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1761	110.93	5.23	5.80	106 mg/dL	-0.85	0.28	

SDI Range	Interpretation				
Within -1.00 to +1.00	Excellent.				
Within ±1.01 to ±2.00	Good.				
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.				
Beyond ±3.0	Unacceptable performance. Action Signal.				

LAB ADDRESS:

Govt. PHC, Pasu Mathur PASU MATHUR Pasu Mathur TAMILNADU635083

Panela Christudoss

Coordinator Contact Details: Email:clinqc@cmcvellore.ac.in Contact Number: 0416-2283102 Dr. Pamela Christudoss CMC EQAS Coordinator Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

******** End of Report *******