



CALIBRATION REPORT	
Report No: HBMS/CAL/23-24	Calibration Date: 23.09.2023 Calibration Valid: 22.03.2024
<b>1.1. CUSTOMER DETAILS:</b>	
Name and Address of the organization	M/S. CORPORATION HOSPITAL, UPHC, PODANUR , COIMBATORE.
<b>1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):</b>	
EQUIPMENT NAME	PIPETTE
I Manufactured by	MICROLUX
II Model	5µl - 50µl
III Serial No	CH/PD/PIP/001
IV Location	LABORATORY


**Calibration standards used :**

Master instrument	Model	Serial no	Date of calibration	Calibration due on
Electrical compact scale	SF-400C	SF400-2240	07-OCT-2022	06-OCT-2023

**Details of the machine:** machine calibrated and checked output.

Number of measurement	Channel 1 (Volume)		
	05µl	20µl	50µl
1	05.01	20.03	50.05
2	05.03	20.04	50.02
3	05.01	20.05	50.04
4	05.03	20.02	50.02
5	05.01	20.06	50.05
6	05.02	20.02	50.04

For Heart Beat Medical Systems,

  
(Authorized signatory)

