

| | | | CALIE | BRAT | TION | CERT | FICAT | <u> </u> | |
|---|---|---|---|-------------------------------------|--|------------------------------|----------------------------------|-----------------------|-------------------|
| CERTI | FICATE NO: | SBS/CL/23 | /12094 | | | | MECHANIC | AL | Page No:1 of 1 |
| ssue D | | | | | 30-09-202 | 23 | | | |
| | o & Date | | | | SRF/23/0 | 0477-0001 | & 29-09-2023 | | |
| Receipt Date | | | | | 29-09-202 | | | | |
| Calibration Date | | | | | 29-09-202 | 23 | | | |
| | tion Due | | | | 28-09-202 | 24 | | | |
| | mer Name & Addr | ess | | | | | | | |
| | RNMENT PRIMAR | | CENTRE, | | | | | | |
| KEELA | THOOVAL-62370 | 6,RAMANAT | HAPURAM [| DISTRIC | Г. | | | | |
| | | | Deta | ails of De | evice Und | ler Calibrat | ion (DUC) | | |
| Description : CENTRIFUGE | | | | | Make & Model MICROSIL & MK-10 | | | | |
| | | 5500 | | Sr. No | | | 279/385 | | |
| Range Resolu | | | RPM | | Identifica | tion No | | NA | |
| | | | | | Location | | | LABORATORY | |
| | Condition : | Satisfactor | | onditior | | dard Opera | ting Procedu | | |
| Enviro | nmental Details | | Temperature | | | Relative Hu | | 55% Rh | |
| | ation Procedure No | | | | | Calibration done at | | ONSITE | |
| JailDie | alion Procedure No | | ODO/OT /IVIE/ | | | ndards Det | | | |
| S.No | Description | | | Make/ S | | | Certificate | No | Validity |
| 5.140 | | | | LINE SEIKI / 175-0034V | | 000.01 | JRPM-CCTR-A&S-2023-001 | | 09-06-2024 |
| 1 | Digital Tachomete | er | | LINE SI | =IKI / 1/5- | -0034V | JRPM-CCT | R-A&S-2023-0013 | 09-06-2024 |
| CALIE | RATION RESULT | S | | | | | | | |
| S.No | DEVICE UNDER STANDARD INS | | | | RUMENTS DEVIATION E | | EXPANDED | EXPANDED UNCERTAINTY | |
| 5.140 | CALIBRATION | | | | | | | 0/ | |
| | RPM | | RPM | | | RPM | | % | |
| 1 | 1000 | | 999.2 | | | 0.8 | | 5.7 | |
| 2 | 2000 | | | 999.6 | | | 0.4 | | 5.7 |
| 3 | 3000 | | | 2999.7 | | | 0.3 | | 5.7 |
| REMA | Calibration certific user should deterr | nine the suit | ability of the i | | | | | the laboratory. | |
| 2.The 3.The 4.The 5.The k=2.0 | 0. | nis certificate nties are exp | relate only to anded uncer | on the us the item tainty est | ser require a calibrate imated for | ement. d. r a confiden | ce level of ap | proximately 95% for a | coverage factor |
| 2.The 3.The 4.The 5.The k=2.0 | results stated in th indicated uncertain | iis certificate nties are exp alibration we | relate only to anded uncer ere calibrated | on the us the item tainty est | ser require a calibrate imated for | ement. d. r a confiden | ce level of ap rnational Star | | i coverage factor |