



CALIBRATION REPORT	
Report No : HBMS/CAL/23-24	Calibration Date: 23.09.2023 Calibration Valid: 22.03.2024
1.1.CUSTOMER DETAILS:	
Name and Address of the organization	M/S. CORPORATION HOSPITAL, UPHC, PODANUR , COIMBATORE.
1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):	
EQUIPMENT NAME	MICRO PIPETTE
i Manufactured by	THERMO SCIENTIFIC
ii Model	FINEPITTE F3
iii Value	100µl-1000µl
iv Serial No	PW02589

Calibration standards used :

Master instrument	Model	Serial no	Date of calibration	Calibration due on
Electrical compact scale	SF-400C	SF400-2240	07-OCT-2022	06-OCT-2023

Calibration test Report.

Number of measurement	Channel 1 (Volume)		
	100µl	500µl	1000µl
1	100.02	500.06	1000.01
2	100.03	500.05	1000.03
3	100.05	500.02	1000.01
4	100.02	500.10	1000.08
5	100.01	500.02	1000.04
6	100.05	500.07	1000.02

For Heart Beat Medical Systems,


(Authorized signatory)

