



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - APRIL 2023



Lab Name: Govt. PHC, Somasipadi Lab No: 12926
 Constituent Group: Chemistry III Date of Result Entered: 20/04/2023
 PT Item: Lyophilized human serum based Date of Report Published: 06/05/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	CV	SD
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1832	119.45	6.04	7.21	121 mg/dL	0.21	0.34
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1820	2.16	8.09	0.18	2.5 mg/dL	1.94	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1750	111.61	5.19	5.79	113 mg/dL	0.24	0.28

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :
 Govt. PHC, Somasipadi
 Somasipadi
 TAMILNADU

Coordinator Contact Details:
 Email: clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pamela Christudoss
 Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

S.M. Dh
 சி.மீ.க.க. வெலூர்
 சி.மீ.க.க. வெலூர்
 தொலைபேசி - 606 611

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential