



AVU-2023

# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - AUGUST 2023



Lab Name: Govt. UPHC, Perandipalayam  
Constituent Group: Chemistry III  
PT Item: Lyophilized human serum based

Lab No: 10863  
Date of Result Entered: 19/08/2023  
Date of Report Published: 01/09/2023

SL No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1841	175.58	5.13	9.00	169.2 mg/dL	-0.71	0.42
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1852	1.76	6.77	0.12	1.95 mg/dL	1.60	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1764	113.80	4.07	4.64	115 mg/dL	0.26	0.22
SDI Range				Interpretation						
Within -1.00 to +1.00				Excellent.						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

LAB ADDRESS :  
Govt. UPHC, Perandipalayam  
PERIYANDIPALAYAM  
Tirupur  
TAMILNADU641687

Coordinator Contact Details:  
Email: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)  
Contact Number: 0416-2283102

*Parula Christudass*  
Dr. Pamela Christudass  
CMC EQAS Coordinator  
Christian Medical College, Vellore

MEDICAL OFFICER  
UPHC - Perandipalayam