

VIEW LAB MONTHLY SUMMARY

Lab Name
 Lab No
 Month
 Year
 Constituent Group

URBAN PRIMARY
 HEALTH CENTER
 10681
 January
 2021
 Chemistry III
 Click on the
 analyte to view
 Graphical Data

[Details About
 Robust Analysis
 Monthly](#)
[Detail about SDI](#)

Date of Result Entered
 : 19/01/2021
 Date of Report Published
 : 05/02/2021

SI No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Any Analyser (Automation / Semi Automation)	117	109.82	7.56	8.3	119 mg/dl	1.11	1.53
2	UREA III	Others (Urease Berthelot / Ned Dye) III	Any Analyser (Automation / Semi Automation)	68	34.8	8.79	3.06	34 mg/dl	-0.26	0.74
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Any Analyser (Automation / Semi Automation)	113	1.35	6.61	0.09	1.6 mg/dl	2.81	0.02
4	CHOLESTROL III	CHOD-PAP III	Any Analyser (Automation / Semi Automation)	118	110.08	6.97	7.68	124 mg/dl	1.81	1.41
SDI Range		Interpretation								
Within -1.0 to +1.0		Excellent.								
Between ±1.0 to ±2.0		Good.								
Between ±2.0 to ±3.0		Accept with caution. Warning Signal.								
Beyond ±3.0		Unacceptable performance. Action Signal.								

M
 MEDICAL OFFICER
 PEELAMEDU UPHC,
 PEELAMEDU,
 COIMBATORE - 4