

VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. PHC, Padalur										
Lab No	12491										Details About Robust Analysis
Year	2023										Detail About Monthly Summary
Month	July										Detail about SDI
Constituent Group	Chemistry III										
All Method Result			Print								
Result Entered Date : 17/07/2023											
Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U	
						CV	SD				
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1876	109.35	5.55	6.07	129.3 mg/dL	3.29	0.28	
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1850	5.83	6.00	0.35	6.15 mg/dL	0.91	0.02	
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1776	113.00	4.12	4.66	117.9 mg/dL	1.05	0.22	
SDI Range		Interpretation									
Within -1.00 to +1.00		Excellent.									
Within ±1.01 to ±2.00		Good.									
Within ±2.01 to ±2.99		Accept with caution. Warning Signal.									
Beyond ±3.0		Unacceptable performance. Action Signal.									

Links

[Modify Address](#)[Modify Methods](#)[Submit Lab Results](#)[View Lab Monthly Summary](#)[Yearly Summary](#)[Home](#)

