

VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. PHC, Padalur									
Lab No	12491									Details About Robust Analysis
Year	2023									Detail About Monthly Summary
Month	June									Detail about SDI
Constituent Group	Chemistry III									
All Method Result			Print							
Result Entered Date : 20/06/2023										
Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1823	286.63	5.17	14.82	286.78 mg/dL	0.01	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1848	2.57	6.64	0.17	2.6 mg/dL	0.18	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1804	113.73	4.46	5.07	111.66 mg/dL	-0.41	0.24
SDI Range		Interpretation								
Within -1.00 to +1.00		Excellent.								
Within ±1.01 to ±2.00		Good.								
Within ±2.01 to ±2.99		Accept with caution. Warning Signal.								
Beyond ±3.0		Unacceptable performance. Action Signal.								

Links

[Modify Address](#)[Modify Methods](#)[Submit Lab Results](#)[View Lab Monthly Summary](#)[Yearly Summary](#)[Home](#)

