

## VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. PHC, Padalur										
Lab No	12491								<a href="#">Details About Robust Analysis</a>		
Year	2023								<a href="#">Detail About Monthly Summary</a>		
Month	May								<a href="#">Detail about SDI</a>		
Constituent Group	Chemistry III										
<a href="#">All Method Result</a>				<a href="#">Print</a>							
Result Entered Date : 19/05/2023											
Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U	
						CV	SD				
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1834	180.88	5.19	9.39	159 mg/dL	-2.33	0.44	
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1854	1.76	7.48	0.13	1.75 mg/dL	-0.08	0.01	
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1798	114.60	5.08	5.83	113.4 mg/dL	-0.21	0.27	
<b>SDI Range</b>		<b>Interpretation</b>									
Within -1.00 to +1.00		Excellent.									
Within ±1.01 to ±2.00		Good.									
Within ±2.01 to ±2.99		Accept with caution. Warning Signal.									
Beyond ±3.0		Unacceptable performance. Action Signal.									

## Links

[Modify Address](#)[Modify Methods](#)[Submit Lab Results](#)[View Lab Monthly Summary](#)[Yearly Summary](#)[Home](#)

