

CHRISTIAN MEDICAL COLLEGE
 DEPARTMENT OF CLINICAL BIOCHEMISTRY
 CMC EXTERNAL QUALITY ASSURANCE SCHEME
 MONTHLY SUMMARY REPORT - MARCH 2023



Lab Name: Govt. PHC, Su. Valavetti
 Constituent Group: Chemistry III
 PT Item: Lymphilized human serum based
 Lab No: 12943
 Date of Result Entered: 20/03/2023
 Date of Report Published: 04/04/2023

SI.No	Analyte	Method / Principle	Analyzer Name	No of Participants	DV	CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1878	91.48	7.46	6.82	90.5	-0.14	0.31
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1851	1.33	9.08	0.12	1.45	-0.99	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1761	110.93	5.23	5.80	115.3	0.75	0.28

SDI Range		Interpretation
Within -1.00 to +1.00	Excellent	
Within ±1.01 to ±2.00	Good	
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.	
Beyond ±3.0	Unacceptable performance. Action Signal.	

LAB ADDRESS :
 Govt. PHC, Su. Valavetti
 Su. VALAVETTI VILLAGE
 Su.Valavetti
 TAMILNADU606806

Coordinator Contact Details:
 Email:clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pamela Christudoss
 Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential
 CMC EQAS does not sub contract any components
 ***** End of Report *****

K. Govardhan
 Director
 Christian Medical College, Vellore
 0416-2283102