



CALIBRATION CERTIFICATE

Certificate No: SBS/CL/23/11773

Page No : 1 of 1

Customer Name & Address

GOVERNMENT PRIMARY HEALTH CENTRE, KALPADI-621113.	SRF No. SRF Date Date of Receipt Date of Calibration Due Date for Calibration Issue Date	SRF/23/00411-0001 26-09-2023 25-09-2023 26-09-2023 25-09-2024 27-09-2023
--	---	---

Details of Unit Under Calibration

Description	MICRO PIPETTE	Make	MICROLUX
Range	5-50µl	Model	NA
Resolution	1µl	Material	PVC
Serial Number	NA	Operating Range	5-50µl
ID Number	NA	Condition of UUC	Good
Cal. At	Mechanical Lab	Instrument Location	LABORATORY

Environmental Condition				Calibration Method Used	
Temperature (°C)	25.7	Humidity (%RH)	53	National / International Standard	ISO 8655-6:2002
Atmospheric Pressure (mbar)	1006	Water Temperature (°C)	21.6	Cal Procedure No	SBS/CP/ML/08

Standard Used

SI. No.	Description	ID.No. / SI. No.	Certificate No.	Make/Model	Traceability	Valid till
1	Electronic Weighing Balance	15112918	TVCSPL22/12/2115-01	A&D & GH-252	National Standards	09-12-2023

Z Factor: 1.00319

Result of Calibration in µl										
Sl. No.	Nominal Value	Observed Readings					Mean Value	Systematic Error	Random Error	Measurement Uncertainty (±)
1	10	9.89	9.88	9.87	9.88	9.88	9.88	-0.12	0.01	0.47
		9.88	9.86	9.87	9.88	9.89				
2	30	29.76	29.77	29.78	29.79	29.76	29.78	-0.22	0.01	0.47
		29.77	29.78	29.79	29.78	29.78				
3	50	49.88	49.87	49.88	49.87	49.88	49.87	-0.13	0.01	0.47
		49.87	49.86	49.86	49.87	49.88				

Remarks

1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
2. The user should determine the suitability of the instrument for its intended use.
3. The recalibration interval should be determined on the user requirement.
4. The results stated in this certificate relate only to the item calibrated.
5. Equipment used for Calibration were calibrated & traceable to National & International Standards
6. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00 .
7. Calibration Liquid Used: Distilled or Deionized water conforming Grade 3 as specified in ISO 3696.

Calibrated By,

(Calibration Engineer)
M.RAGUL



Authorised by:

(Quality Manager/Chief Executive)
C.SIVABALAN

SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,
E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.



CALIBRATION CERTIFICATE

Certificate No: SBS/CL/23/11774

Page. No : 1 of 1

Customer Name & Address

GOVERNMENT PRIMARY HEALTH CENTRE, KALPADI-621113.	SRF No.	SRF/23/00411-0002
	SRF Date	26-09-2023
	Date of Receipt	25-09-2023
	Date of Calibration	26-09-2023
	Due Date for Calibration	25-09-2024
	Issue Date	27-09-2023

Details of Unit Under Calibration

Description	MICRO PIPETTE	Make	MICROLUX
Range	100-1000µl	Model	NA
Resolution	1µl	Material	PVC
Serial Number	NA	Operating Range	100-1000µl
ID Number	NA	Condition of UUC	Good
Cal. At	Mechanical Lab	Instrument Location	LABORATORY

Environmental Condition				Calibration Method Used	
Temperature (°C)	25.7	Humidity (%RH)	53	National / International Standard	ISO 8655-6:2002
Atmospheric Pressure (mbar)	1006	Water Temperature (°C)	21.6	Cal Procedure No	SBS/CP/ML/08

Standard Used

SI. No.	Description	ID.No. / SI. No.	Certificate No.	Make/Model	Traceability	Valid till
1	Electronic Weighing Balance	15112918	TVCSPL22/12/2115-01	A&D & GH-252	National Standards	09-12-2023

Z Factor: 1.00319

Result of Calibration in µl

Sl. No.	Nominal Value	Observed Readings					Mean Value	Systematic Error	Random Error	Measurement Uncertainty (±)
1	100	99.93	99.94	99.92	99.95	99.96	99.94	-0.06	0.02	0.47
		99.95	99.93	99.96	99.97	99.93				
2	500	499.92	499.95	499.96	499.94	499.96	499.96	-0.04	0.02	0.47
		499.95	499.97	499.98	499.99	499.96				
3	1000	999.95	999.96	999.97	999.98	999.96	999.92	-0.08	0.09	0.47
		999.95	999.98	999.96	999.74	999.75				

Remarks

- 1.This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2.The user should determine the suitability of the instrument for its intended use.
- 3.The recalibration interval should be determined on the user requirement.
- 4.The results stated in this certificate relate only to the item calibrated.
5. Equipment used for Calibration were calibrated & traceable to National & International Standards
- 6.The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00 .
7. Calibration Liquid Used: Distilled or Deionized water conforming Grade 3 as specified in ISO 3696.

Calibrated By,

(Calibration Engineer)
 M.RAGUL



Authorised by:

(Quality Manager/Chief Executive)
 C.SIVABALAN

SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,
E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.