



NEW LAB MONTHLY SUMMARY

Lab Name	Govt. PHC, Thiruvananthapuram	Details About Robust Analysis
Lab NO	12212	Detail About Monthly Summary
Month	August	Detail about SDI
Year	2023	
Constituent Group	Chemistry II	

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#) [Print](#)

Date of Result Entered : 21/08/2023

Date of Report Published : 31/09/2023

Sl. No	Analyte	Method / Principle Name	Analyser Name	No. of Participants	CV	Participants		Your Value	SDI	U
						DV	SI			
1	GLUCOSE W	GOD-POD II	Robovik [Semi automatic / Fully Automatic]	1041	1.73.58	0.12	9.00	181.3 mg/dL	1.33	0.42
2	CREATININE W	Jeffco Kinetic - Alkaline Phosph II	Robovik [Semi automatic / Fully Automatic]	1032	1.76	0.77	0.12	2.1 mg/dL	2.86	0.01
3	CHOLESTROL W	CHOD-PAP II	Robovik [Semi automatic / Fully Automatic]	1764	112.80	4.07	4.64	117 mg/dL	0.83	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within +1.01 to +2.00	Good.
Within +2.01 to +2.99	Accept with caution. Warning Signal.
Beyond +3.0	Unacceptable performance. Action Signal.

Links			
Notify Address	Notify Methods	Submit Lab Results	View Lab Monthly Summary
Notify Summary	Change Password	Home	Logout