

VIEW LAB MONTHLY SUMMARY

| Lab Name | Govt. PHC, Thiruvithancode | Details About Robust Analysis | | | | | | | | | |
|-----------------------|--------------------------------|---|--|--------------------|--------|-------------------------------------|-----------------|-----------------------|-------|------|--|
| Lab No | 12279 | Detail About Monthly Summary | | | | | | | | | |
| Month | June | Detail about SDI | | | | | | | | | |
| Year | 2023 | | | | | | | | | | |
| Constituent Group | Chemistry III | | | | | | | | | | |
| | | Click on the analyte to view Graphical Data | | | | All Analyser Result | | Print | | | |
| | | Date of Result Entered : 21/06/2023 | | | | | | | | | |
| | | Date of Report Published : 04/07/2023 | | | | | | | | | |
| Sl.No | Analyte | Method / Principle Name | Analyzer Name | No of Participants | DV | Participants CV | Participants SD | Your Value | SDI | U | |
| 1 | GLUCOSE III | GOD-POD III | Robonik (Semi automatic / Fully Automatic) | 1823 | 286.63 | 5.17 | 14.82 | 297 mg/dL | 0.70 | 0.69 | |
| 2 | CREATININE III | Jaffes Kinetic - Alkaline Picrate III | Robonik (Semi automatic / Fully Automatic) | 1848 | 2.57 | 6.64 | 0.17 | 2.4 mg/dL | -0.99 | 0.01 | |
| 3 | CHOLESTROL III | CHOD-PAP III | Robonik (Semi automatic / Fully Automatic) | 1804 | 113.73 | 4.46 | 5.07 | 112 mg/dL | -0.34 | 0.24 | |
| SDI Range | | Interpretation | | | | | | | | | |
| Within -1.00 to +1.00 | | Excellent. | | | | | | | | | |
| Within ±1.01 to ±2.00 | | Good. | | | | | | | | | |
| Within ±2.01 to ±2.99 | | Accept with caution. Warning Signal. | | | | | | | | | |
| Beyond ±3.0 | | Unacceptable performance. Action Signal. | | | | | | | | | |

Links

| | | | |
|--------------------------------|---------------------------------|------------------------------------|--|
| Modify Address | Modify Methods | Submit Lab Results | View Lab Monthly Summary |
| Yearly Summary | Change Password | Home | Sign-out |