

VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. PHC, Thiruvithancode		Details About Robust Analysis							
Lab No	12279		Detail About Monthly Summary							
Month	May		Detail about SDI							
Year	2023									
Constituent Group	Chemistry III									
Click on the analyte to view Graphical Data			All Analyser Result		Print					
			Date of Result Entered : 19/05/2023							
			Date of Report Published : 05/06/2023							
Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	Participants SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1834	180.88	5.19	9.39	167 mg/dL	-1.48	0.44
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1854	1.76	7.48	0.13	1.9 mg/dL	1.06	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1798	114.60	5.08	5.83	112 mg/dL	-0.45	0.27
SDI Range			Interpretation							
Within -1.00 to +1.00			Excellent.							
Within ±1.01 to ±2.00			Good.							
Within ±2.01 to ±2.99			Accept with caution. Warning Signal.							
Beyond ±3.0			Unacceptable performance. Action Signal.							

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out