



**CALIBRATION REPORT**

Report No : HBMS/CAL/23-24		Calibration Date : 20.09.2023
		Calibration valid : 19.03.2024
<b>1.1.CUSTOMER DETAILS:</b>		
Name and Address of the organization		<b>M/s, PRIMARY HEALTH CENTRE , PULIYUR, THIRUVANNAMALAI-635703.</b>
<b>1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):</b>		
EQUIPMENT NAME		CENTRIFUGE
<b>i</b>	Manufactured by	REMI
<b>ii</b>	Model	C8C
<b>iii</b>	Serial No	PHC/PU/CENT/001

**CALIBRATION WORKSHEET**

**LOCATION: LABORATORY**

**Tabulation For – Centrifuge**

S.no	Unit under test IN Minutes	Observed speed tachometer in RPM	Deviation measured $\pm$ RPM
1	1000	1000	0
2	1200	1201	1
3	1500	1502	2
4	2000	2002	2

Acceptable deviation is  $\pm$  5 RPM it depends on Max Speed Range



MD Reg No: TN/CBE/MD42/00016

GSTIN: 33AAJFH1833M1ZB



99400 28355, 9751206106

88833 15662



No.42/1, Navaindia Road, Peelamedu,

CBE – 04, hbmedicalsystem@gmail.com