

CHRISTIAN MEDICAL COLLEGE
 DEPARTMENT OF CLINICAL BIOCHEMISTRY
 CMC EXTERNAL QUALITY ASSURANCE SCHEME
 MONTHLY SUMMARY REPORT - MARCH 2023



Lab Name: Govt. PHC, Vathalmalai
 Lab No: 12126
 Constituent Group: Chemistry III
 Date of Result Entered: 10/03/2023
 Test Item: Lyophilized human serum based
 Date of Report Published: 04/04/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1878	91.48	7.46	6.82	104 mg/dL	1.84	0.31
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1851	1.33	9.08	0.12	0.8 mg/dL	-4.42	0.01
3	CHOLESTROL III	CHOD-PAP III	Any Analyser (Automation / Semi Automation)	148	111.55	4.29	4.78	127 mg/dL	3.23	0.79

SDI Range	Interpretation
within -1.00 to +1.00	Excellent.
within ±1.01 to ±2.00	Good.
within ±2.01 to ±2.99	Accept with caution. Warning Signal.
beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :
 Govt. PHC, Vathalmalai
 THE MEDICAL OFFICER, VATHALMALAI PHC, PAPPIREDDIPATTI, DHARMAPURI DIST.
 Vathalmalai
 MILNADU636704

Coordinator Contact Details:
 Email: clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pamela Christudoss
 Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential
 CMC EQAS does not sub contract any components

***** End of Report *****