



CHRISTIAN MEDICAL COLLEGE



DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - MARCH 2023

PC-1024

Lab Name **Govt. PHC, Ladapuram**

Lab No **7129**

Constituent Group **Chemistry III**

Date of Result Entered : **20/03/2023**

PT item **Lyophilized human serum based**

Date of Report Published : **04/04/2023**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1878	91.48	7.46	6.82	91 mg/dL	-0.07	0.31
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1851	1.33	9.08	0.12	1.2 mg/dL	-1.07	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1761	110.93	5.23	5.80	101 mg/dL	-1.71	0.28

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

Govt. PHC, Ladapuram

"The Medical Officer, Ladapuram, Ladapuram Po, Perambalur Tk & Dt.

Ladapuram

TAMILNADU621121

Pamela Christudoss

**Coordinator Contact Details:
Email:clinqc@cmcvellore.ac.in
Contact Number: 0416-
2283102**

**Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College,
Vellore**

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

******* End of Report *******