



CHRISTIAN MEDICAL COLLEGE



DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - APRIL 2023

PC-1024

Lab Name **Govt. PHC, Ladapuram**

Lab No **7129**

Constituent Group **Chemistry III**

Date of Result Entered : **20/04/2023**

PT item **Lyophilized human serum based**

Date of Report Published : **06/05/2023**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1832	119.45	6.04	7.21	126 mg/dL	0.91	0.34
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1820	2.16	8.09	0.18	2.26 mg/dL	0.57	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1750	111.61	5.19	5.79	109 mg/dL	-0.45	0.28

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

Govt. PHC, Ladapuram

"The Medical Officer, Ladapuram, Ladapuram Po, Perambalur Tk & Dt.

Ladapuram

TAMILNADU621121

Pamela Christudoss

**Coordinator Contact Details:
Email:clinqc@cmcvellore.ac.in
Contact Number: 0416-
2283102**

**Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College,
Vellore**

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

******* End of Report *******