



DR.C.VELRAJAN

SENIOR ASSISTANT SURGEON

MEDICAL OFFICER

MELAKADAYANALLUR.

OFFICE OF GOVT. OF URBAN PRIMARY HEALTH CENTER

MELAKADAYANALLUR

KADAYANALLUR -627751

TENKASI DISTRICT.

29.09.2023

LEGAL IDENTITTY CERTIFICATE

This is certify that the following Primary Health centre is functioning under the department of Public Health and Preventive Medicine , Government of Tamil Nadu.

| Name of the District | Name of the Health Unit District | Name of the Block | Name of the Primary Health Centre |
|----------------------|----------------------------------|-------------------|-----------------------------------|
| Tenkasi | Tenkasi | Chokkampatti | Mela Kadayanallur Urban |

This certificate is issued for submitting the Application for the "NABL Medical Entry Level Testing M(EL)T Labs program"


MEDICAL OFFICER
URBAN PRIMARY HEALTH CENTRE
KADAYANALLUR - 627 751



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - AUGUST 2023



PC-1024

Lab Name: Govt. UPHC, Mela Kadayanallur
Constituent Group: Chemistry III
PT Item: Lyophilized human serum based

Lab No: 10879
Date of Result Entered: 19/08/2023
Date of Report Published: 01/09/2023

| Sl.No | Analyte | Method / Principle Name | Analyzer Name | No of Participants | DV | CV | SD | Your Value | SDI | U |
|-----------------------|----------------|---------------------------------------|--|--|--------|-------|------|------------|-------|------|
| 1 | GLUCOSE III | GOD-POD III | Robonik (Semi automatic / Fully Automatic) | 1841 | 175.58 | 5.13 | 9.00 | 185 mg/dL | 1.05 | 0.42 |
| 2 | UREA III | Others (Urease Berthelot / Ned Dye) | Any Analyser (Automation / Semi Automation) | 59 | 28.92 | 13.99 | 4.05 | 34 mg/dL | 1.26 | 1.05 |
| 3 | CREATININE III | Jaffes Kinetic - Alkaline Picrate III | Robonik (Semi automatic / Fully Automatic) | 1852 | 1.76 | 6.77 | 0.12 | 1.7 mg/dL | -0.50 | 0.01 |
| 4 | CHOLESTROL III | CHOD-PAP III | Robonik (Semi automatic / Fully Automatic) | 1764 | 113.80 | 4.07 | 4.64 | 110 mg/dL | -0.82 | 0.22 |
| SDI Range | | | | Interpretation | | | | | | |
| Within -1.00 to +1.00 | | | | Excellent. | | | | | | |
| Within ±1.01 to ±2.00 | | | | Good. | | | | | | |
| Within ±2.01 to ±2.99 | | | | Accept with caution. Warning Signal. | | | | | | |
| Beyond ±3.0 | | | | Unacceptable performance. Action Signal. | | | | | | |

LAB ADDRESS :
Govt. UPHC, Mela Kadayanallur
MELA KADAYANALLUR
Kadayanallur
TAMILNADU627751

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss

Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

C. Helen
30/9/23
MEDICAL OFFICER
URBAN PRIMARY HEALTH CENTRE
MADAYANALLUR 627 751

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****



ROBONIKTM

prietest

TOUCH

Auto-Biochemistry Analyzer

PRIESTEST TOUCH 30/9/23 35.00
 MENU
 1 GLUCOSE
 2 CHOL
 3 HB
 4 UREA-BER
 5 UREA-K
 6 CREATINI
 7 TGL
 8 HDL
 9 BILI-T
 10 BILI-D
 PRNT
 NXT
 PREV
 DEL
 EDIT



FEED

PRINT

WASH

YES

NO

ESC



MELA KADAYANALLUR
UPHC




 Call - 93664 33099
 SI No. **AT 246/11/10**
 Calibration Date On **01/02/2023**
 Valid Upto **01/02/2024**
Orion Diagnostics
 400 A, Maniarman Kovil Street, Srivilliputhur - 620125