



# CHRISTIAN MEDICAL COLLEGE



## DEPARTMENT OF CLINICAL BIOCHEMISTRY CMC EXTERNAL QUALITY ASSURANCE SCHEME

PC-1024

### MONTHLY SUMMARY REPORT - MAY 2023

Lab Name **Govt. PHC, Ladapuram**

Lab No

**7129**

Constituent Group **Chemistry III**

Date of Result Entered :

**18/05/2023**

PT item **Lyophilized human serum based**

Date of Report Published :

**05/06/2023**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1834	180.88	5.19	9.39	<b>169</b> mg/dL	-1.26	0.44
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1854	1.76	7.48	0.13	<b>1.56</b> mg/dL	-1.52	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1798	114.60	5.08	5.83	<b>114</b> mg/dL	-0.10	0.27

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

**LAB ADDRESS :**

**Govt. PHC, Ladapuram**

**"The Medical Officer, Ladapuram, Ladapuram Po, Perambalur Tk & Dt.**

**Ladapuram**

**TAMILNADU621121**

*Pamela Christudoss*

**Coordinator Contact Details:  
Email:clinqc@cmcvellore.ac.in  
Contact Number: 0416-  
2283102**

**Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College,  
Vellore**

**Homogeneity and Stability of the sample is passed.**

**Data in CMC EQAS reports is confidential**

**CMC EQAS does not sub contract any components**

**\*\*\*\*\* End of Report \*\*\*\*\***