



Certificate No : GPHCT/005

**BIOMEDICAL
TECHNIQUES**
Calibration experts**CALIBRATION CERTIFICATE**

Customer Details	Details of DUC	
GOVT PRIMARY HEALTH CENTRE, KAMALAPANTHY STREET, THIRUVITHANCODE, TAMILNADU-629174	Description	CENTRIFUGE
	Manufactured by	NIL
	Model	NIL
	Serial No.	NIL
	Identification No.	NIL
	Location	LABORATORY

Environmental Conditions		Calibration Details	
Operating Temperature	24°C	Date of Receipt	13/12/2022
Humidity	63 %	Date of Calibration	13/12/2022
		Date of Issue	15/12/2022
		Due Date of calibration	12/12/2023
		Calibrated At	Site

Details of Calibration Standards

Name	Manufacturer	Serial No.	Valid up to	Traceability
Electrical safety analyser	Fluke / ESA 615	2244039	26/09/2023	Accuver calibration laboratory co llc
Digital Tacho Meter	Amprobe	BMT TACHO 031	30/09/2023	Equations,Chennai

Performance Analysis

Parameter	Set Value(RPM)	Measured Value(RPM)	Tolerance	Error Observed (\pm RPM)
RPM	1500	1502	$\pm 5\%$	2
	2000	2008	$\pm 5\%$	8

+91-487-24 25 821
+91-98 95 02 58 21
+91- 90 48 44 51 11

biomedicaltechniques@gmail.com

Registered Office : II/446, S.N. Park, Poothole P.O.
Thrissur, Kerala, India - 680 004Corporate Office : XXV/277, Unity Building
Mannadiar Lane, Thrissur, Kerala - 680001

www.bmtqa.com

QUALITY MANAGEMENT ISO 9001 : 2008

செயல்பாடு வாய்மைக்கு! புறம் தாய்மை வாழ்வுக்கு!