



No : GPHCT/007

**BIOMEDICAL
TECHNIQUES**
Calibration experts

CALIBRATION CERTIFICATE

Customer Details	Details of DUC	
GOVT PRIMARY HEALTH CENTRE, KAMALAPANTHY STREET, THIRUVITHANCODE, TAMILNADU-629174	Description	BIOCHEMISTRY ANALYZER
	Manufactured by	ROBONIK
	Model	PRIETEST TOUCH
	Serial No.	ATCD4060321RBK
	Identification No.	NIL
	Location	LABORATORY

Environmental Conditions		Calibration Details	
Operating Temperature	24°C	Date of Receipt	13/12/2022
		Date of Calibration	13/12/2022
Humidity	63 %	Date of Issue	15/12/2022
		Due Date of calibration	12/12/2023
		Calibrated At	Site

Details of Calibration Standards

Name	Manufacturer	Serial No.	Valid up to	Traceability
Electrical safety analyser	Fluke / ESA 615	2244039	26/09/2023	Accuver calibration laboratory co llc

Electrical Safety Test

Parameter	Measured Value	Tolerance	Remarks
Ground wire Resistance	0.184 Ω	<0.3 Ω	pass
insulation Resistance	OL	> 2 M Ω	pass
Chassis Leakage Current *NC	0.233 μ A	<100 μ A	Pass
Chassis Leakage Current *SFC	0.50 μ A	<500 μ A	Pass
Patient Leakage Current - B	0.81 μ A	<100 μ A	pass