



CHRISTIAN MEDICAL COLLEGE
 DEPARTMENT OF CLINICAL BIOCHEMISTRY
 CMC EXTERNAL QUALITY ASSURANCE SCHEME
 MONTHLY SUMMARY REPORT - AUGUST 2023



Lab Name: Govt. PHC, Mungilthuraiipattu

Lab No: 13120

Constituent Group: Chemistry III

Date of Result Entered: 19/08/2023

PT Item: Lympholized human serum based

Date of Report Published: 01/09/2023

Sl. No.	Analysis	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1841	175.58	5.13	9.00	179.2 mg/dL	0.40	0.42
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.76	6.77	0.12	1.8 mg/dL	0.34	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1764	113.80	4.07	4.64	112 mg/dL	-0.39	0.22
SDI Range				Interpretation						
Within -1.00 to +1.00				Excellent.						
Within +1.81 to +2.00				Good.						
Within +2.01 to +2.99				Accept with caution. Warning Signal.						
Beyond +3.0				Unacceptable performance. Action Signal.						

LAB ADDRESS :
 Govt PHC, Mungilthuraiipattu
 Mungilthuraiipattu
 TAMILNADU

Coordinator Contact Details:
 Email: dtingdc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Paula Christudoss
 Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

[Signature]
MEDICAL OFFICER
 Govt. Primary Health Centre
 Mungilthuraiipattu-605 103
 Kallakurichi District

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential

