



External Quality Assurance Scheme - Print Monthly Summary

CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
 CMC EXTERNAL QUALITY ASSURANCE SCHEME
 MONTHLY SUMMARY REPORT - JUNE 2023



Lab Name

Govt. PHC, Mungilthuraiipattu

Lab No

13120

Constituent Group

Chemistry III

Date of Result Entered :

20/06/2023

PT Item

Lyophilized human serum based

Date of Report Published :

04/07/2023

SLNo	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	CV	Participants SD	Your Value	SD	%
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1823	286.63	5.17	14.82	279 mg/dL	-0.51	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1848	2.57	6.64	0.17	2.7 mg/dL	0.76	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1804	113.73	4.46	5.07	118.1 mg/dL	0.86	0.24
SDI Range				Interpretation						
Within -1.00 to +1.00				Excellent.						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

LAB ADDRESS :
 Govt PHC, Mungilthuraiipattu
 Mungilthuraiipattu
 TAMILNADU

Coordinator Contact Details:
 Email: cllngqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pamela Christodoss
 Dr. Pamela Christodoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

[Signature]
MEDICAL OFFICER
 Govt Primary Health Centre
 Mungilthuraiipattu-605 702
 Kallakurichi District

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential