



MUTHU SURGICALS

101ANNAI COMPLEX, NEW BUS
 STSND, PERAMBALUR -621212
 GST NO : 33DDEPM1487D1ZQ
 UDYAM REG NO : TN-16-000072
 PHONE NO : 8056356534

CALIBRATION CERTIFICATE

Certificate No: MS16MPT072023				Page. No		
Customer Name & Address						
THE MEDICAL OFFICER, PRIMARY HEALTHCENTRE, LADAPURAM-621101,				Date of Receipt	08-02-2023	
				Date of Calibration	08-02-2023	
				Due Date for Calibration	09-02-2023	
				Issue Date	09-02-2023	
Details Of Unit Under Calibration						
Description	Micro Pipette	Make	THERMO FISHER SCIENTIFIC			
Range	100GL-1000GL	Model	FINNPIPETTE F3			
Resolution		Material	PVC			
Serial Number	RW02887	Operating Range				
ID Number		Condition of UUC				
Cal. At	Mechanical Lab	Instrument Location				
Environmental Condition			Calibration Method Used			
Temperature (°c)	23.9	Humidity (%RH)	55	National / International Standard	ISO 8655-6:2002	
Atmospheric Pressure (mbar)	1006	Water Temperature (°C)	21.6	Cal Procedure No	SBS/CP/ML/08	
Standard Used						
Sl. No.	Description	ID.No. / SL No.	Certificate No.	Make/Model	Traceability	Valid till
1	Electronic Semi Micro Balance	15112918	TVCSP21/12/1587-01	& GH-252	National Standards	09-02-2023