



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - APRIL 2023



PC-1024

Lab Name: Govt. PHC, Mungilthuraiipattu
Constituent Group: Chemistry III
PT Item: Lyophilized human serum based

Lab No: 13120
Date of Result Entered: 20/04/2023
Date of Report Published: 06/05/2023

SI No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value mg/dL	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1832	119.45	6.04	7.21	121 mg/dL	0.21	0.34
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1820	2.16	8.09	0.18	1.9 mg/dL	-1.49	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1750	111.61	5.19	5.79	115.2 mg/dL	0.62	0.28
SDI Range Interpretation										
Within -1.00 to +1.00				Excellent.						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

LAB ADDRESS :
Govt. PHC, Mungilthuraiipattu
Mungilthuraiipattu
TAMILNADU

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Paula Christudoss
Dr. Pamela Christudoss
CMC EOAS Coordinator
Christian Medical College, Vellore

[Signature]
MEDICAL OFFICER
Govt Primary Health Centre
Moongilthuraiipattu-605 702
Kallakurichi District

Homogeneity and Stability of the sample is passed.
Data in CMC EOAS reports is confidential