



**CHRISTIAN MEDICAL COLLEGE**  
 DEPARTMENT OF CLINICAL BIOCHEMISTRY  
 CMC EXTERNAL QUALITY ASSURANCE SCHEME  
 MONTHLY SUMMARY REPORT - JUNE 2023



Lab Name: Govt. UPHC, K.K.Pudu  
 Constituent Group: Chemistry III  
 PT item: Lyophilized human serum based

Lab No: 10851  
 Date of Result Entered: 16/06/2023  
 Date of Report Published: 04/07/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Any Analyser (Automation / Semi Automation)	65	287.55	4.50	12.95	345 mg/dL	4.44	3.21
2	UREA III	Others ( Urease Berthelot / Ncd Dye ) III	Any Analyser (Automation / Semi Automation)	50	33.99	14.82	5.04	28 mg/dL	-1.19	1.42
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Any Analyser (Automation / Semi Automation)	67	2.58	6.39	0.16	2.6 mg/dL	0.12	0.04
4	CHOLESTROL III	CHOD-PAP III	Any Analyser (Automation / Semi Automation)	149	114.61	4.05	4.64	105 mg/dL	-2.07	0.76
<b>SDI Range</b>										
				<b>Interpretation</b>						
Within -1.00 to +1.00				Excellent.						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

**LAB ADDRESS :**  
 Govt. UPHC, K.K.Pudu  
 MEDICAL OFFICER, VI<sup>th</sup> STREET, K.K.PUDUR, SAIBABA COLONY  
 Coimbatore  
 TAMILNADU641038

**Coordinator Contact Details:**  
 Email: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)  
 Contact Number: 0416-2283102

*Pamela Christodoss*  
 Dr. Pamela Christodoss  
 CMC EQAS Coordinator  
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
 Data in CMC EQAS reports is confidential