



CHRISTIAN MEDICAL COLLEGE
 DEPARTMENT OF CLINICAL BIOCHEMISTRY
 CMC EXTERNAL QUALITY ASSURANCE SCHEME
 MONTHLY SUMMARY REPORT - MAY 2023

MONTHLY SUMMARY REPORT - MAY 2023



PC-1024

Lab Name: Govt. UGPHC, Sivagiri

Lab No: 7382

Constituent Group: Chemistry III

Date of Result Entered: 22/05/2023

PT Item: Lyophilized human serum based

Date of Report Published: 05/06/2023

Item No	Test Name	Manufacturer	Result	Unit	Target Range	Acceptance Criteria
1	GLUCOSE III	GOD-POD III	1834	mg/dL	180.88	Excellent.
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	1854	mg/dL	1.76	Good.
3	CHOLESTROL III	CHOD-PAP III	1798	mg/dL	114.60	Accept with caution. Warning Signal.
SDI Range			Interpretation			
Within -1.00 to +1.00			Excellent.			
Within ±1.01 to ±2.00			Good.			
Within ±2.01 to ±2.99			Accept with caution. Warning Signal.			
Beyond ±3.0			Unacceptable performance. Action Signal.			

LAB ADDRESS :

Govt. UGPHC, Sivagiri
 THE BLOCK MEDICAL OFFICER, CLINICAL LABORATORY, ERODE DISTRICT.
 Sivagiri
 TAMILNADU638109

Coordinator Contact Details:
 Email: cilingqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Pamela Christudoss

[Signature]

MEDICAL OFFICER
 GOVT. COMMUNITY HEALTH CENTRE
 SIVAGIRI-639 109, ERODE

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential
 CMC EQAS does not sub contract any components
 ***** End of Report *****