



CHRISTIAN MEDICAL COLLEGE
DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - AUGUST 2023



PC-1024

Lab Name: Govt. UPHC, K.K.Pudu
Constituent Group: Chemistry III

Lab No: 10851

PT item: Lyophilized human serum based

Date of Result Entered: 19/08/2023

Date of Report Published: 01/09/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Any Analyser (Automation / Semi Automation)	64	174.34	5.46	9.51	179 mg/dL	0.49	2.38
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Any Analyser (Automation / Semi Automation)	62	1.76	7.18	0.13	0.9 mg/dL	-6.62	0.03
3	CHOLESTROL III	CHOD-PAP III	Any Analyser (Automation / Semi Automation)	149	113.96	4.36	4.97	112 mg/dL	-0.39	0.81
SDI Range Interpretation										
Within -1.00 to +1.00				Excellent.						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

LAB ADDRESS :

Govt. UPHC, K.K.Pudu
MEDICAL OFFICER, VI th STREET, K.K.PUDUR, SAIBABA COLONY
Coimbatore
TAMILNADU641038

Coordinator Contact Details:

Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christodoss

Dr. Pamela Christodoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

***** End of Report *****