



**CHRISTIAN MEDICAL COLLEGE**  
DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - JULY 2023



Lab Name Govt. PHC, Odathurai

Lab No

12185

Constituent Group Chemistry III

Date of Result Entered :

20/07/2023

PT Item Lyophilized human serum based

Date of Report Published :

07/08/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1876	109.35	5.55	6.07	118.4 mg/dL	1.49	0.28
2	CREATININE III	OTHERS ( Any Other Principles / Methods )	Any Analyser (Automation / Semi Automation)	14	5.89	4.33	0.26	5.7 mg/dL	-0.75	0.14
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1776	113.00	4.12	4.66	122 mg/dL	1.93	0.22
<b>SDI Range Interpretation</b>										
Within -1.00 to +1.00				Excellent.						
Within $\pm 1.01$ to $\pm 2.00$				Good.						
Within $\pm 2.01$ to $\pm 2.99$				Accept with caution. Warning Signal.						
Beyond $\pm 3.0$				Unacceptable performance. Action Signal.						

**LAB ADDRESS :**  
Govt. PHC, Odathurai  
THE MEDICAL OFFICER, BHAVANI  
Odathurai  
TAMILNADU

**Coordinator Contact Details:**  
Email: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)  
Contact Number: 0416-2283102

*Dr. Pamela Christudoss*  
Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

*M. P. S.*  
THE MEDICAL OFFICER  
Govt. Primary Health Centre,  
odathurai, ERODE District - 6  
48

Homogeneity and Stability of the sample is passed.  
Data in CMC EQAS reports is confidential