



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - JUNE 2023



PC-1024

Lab Name Govt. PHC, Odathurai

Lab No 12185

Constituent Group Chemistry III

Date of Result Entered : 20/06/2023

PT Item Lyophilized human serum based

Date of Report Published : 04/07/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1823	286.63	5.17	14.82	280 mg/dL	-0.45	0.69
2	CREATININE III	OTHERS ( Any Other Principles / Methods )	Any Analyser (Automation / Semi Automation )	13	2.55	7.09	0.18	2.2 mg/dL	-1.93	0.10
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1804	113.73	4.46	5.07	110 mg/dL	-0.74	0.24
<b>SDI Range</b>										
				<b>Interpretation</b>						
Within -1.00 to +1.00				Excellent.						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

*M. P. X.*

**LAB ADDRESS :**  
Govt. PHC, Odathurai  
THE MEDICAL OFFICER, BHAVANI  
Odathurai  
TAMILNADU

*Pamela Christodoss*

THE MEDICAL OFFICER  
Govt. Primary Health Centre,  
dathuralesRODE District - 6  
45

**Coordinator Contact Details:**  
Email: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)  
Contact Number: 0416-2283102

Dr. Pamela Christodoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
Data In CMC EQAS reports is confidential