



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - MAY 2023



PC-1024

Lab Name: Govt. PHC, Odathurai
 Constituent Group: Chemistry III
 PT Item: Lyophilized human serum based

Lab No: 12185
 Date of Result Entered: 19/05/2023
 Date of Report Published: 05/06/2023

SL No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1834	180.88	5.19	9.39	171.2 mg/dL	-1.03	0.44
2	CREATININE III	OTHERS (Any Other Principles / Methods)	Any Analyser (Automation / Semi Automation)	15	1.76	8.24	0.14	1.65 mg/dL	-0.76	0.07
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1798	114.60	5.08	5.83	107.8 mg/dL	-1.17	0.27
SDI Range				Interpretation						
Within -1.00 to +1.00				Excellent.						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

LAB ADDRESS :
 Govt. PHC, Odathurai
 THE MEDICAL OFFICER, BHAVANI
 Odathurai
 TAMILNADU

Coordinator Contact Details:
 Email: cllngc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pamela Christudoss
 Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

P. P. P.
 THE MEDICAL OFFICER
 Govt. Primary Health Centre,
 Odathurai, ERODE District - 635 407

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential