



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - APRIL 2023



PC-1024

Lab Name Govt. PHC, Odathural
Constituent Group Chemistry III
PT Item Lyophilized human serum based

Lab No 12185
Date of Result Entered : 20/04/2023
Date of Report Published : 06/05/2023

Sl No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1832	118.45	6.04	7.21	108 mg/dL	-1.59	0.34
2	CREATININE III	OTHERS (Any Other Principles / Methods)	Any Analyser (Automation / Semi Automation)	13	2.20	10.23	0.22	2 mg/dL	-0.89	0.12
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1750	111.61	5.19	5.79	117 mg/dL	0.93	0.28
SDI Range Interpretation										
Within -1.00 to +1.00				Excellent.						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

LAB ADDRESS :
Govt. PHC, Odathural
THE MEDICAL OFFICER, BHAVANI
Odathural
TAMILNADU

Coordinator Contact Details:
Email: cllngc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

M. P. A.
THE MEDICAL OFFICER
Govt. Primary Health Centre,
dathurai.ERODUR District - 6. 4R

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential