

**Dr.B.Bharanidharan, M.B.B.S**  
**and Preventive Medicine**  
**Deputy Director of Health Serv**  
**Chengalpattu Dist-603 001**



**Directorate of Public Health**

**LEGAL IDENTITY CERTIFICATE**

This is to certify that the following Primary Health Center is functioning under the Department of Public Health and Preventive Medicine, Government of Tamil Nadu.

<b>Name of the District</b>	<b>Name of the Health Unit District</b>	<b>Name of the Block</b>	<b>Name of the Primary Health Centre</b>
Chengalpattu	Chengalpattu	Thiruporur	Sembakkam

This certificate is issued for submitting the Application for the "NABL Medical Entry Level Testing M(EL)T Labs Program"

  
**ASSISTANT SURGEON**  
**Govt. Primary Health Centre**  
**Sembakkam-603108.**

**Deputy Director of Health Services,**  
**Chengalpattu.**



Department of Clinical Biochemistry  
Christian Medical College, Vellore - 632004

## CMC EXTERNAL QUALITY ASSURANCE SCHEME

(Under the aegis of ACBI)

### Certificate of Participation

*This is to certify that*

GOVT PRIMARY HEALTH CENTRE

SEMBAKKAM

Lab No. .... 10532 ..... has participated

*in the Chemistry III Program of the External Quality Assurance Scheme  
during the year 2022..*

*Pamela Christudoss*

**Dr. Pamela Christudoss, M.Sc., Ph.D.,**

EOAS Coordinator

Professor of Clinical Biochemistry



Certificate No. PC-1024  
NABL PT Accredited Laboratory

# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - AUGUST 2023



PC-1024

Lab No 10532  
Date of Result Entered : 20/08/2023  
Date of Report Published : 01/09/2023

Sample Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
				CV	SD			
	Robonik ( Semi automatic / Fully Automatic )	1841	175.58	5.13	9.00	177 mg/dL	0.16	0.42
Urea Nitrogen	Robonik ( Semi automatic / Fully Automatic )	1852	1.76	6.77	0.12	1.74 mg/dL	-0.17	0.01
	Robonik ( Semi automatic / Fully Automatic )	1764	113.80	4.07	4.64	107 mg/dL	-1.47	0.22

### Interpretation

Excellent.

Good.

Accept with caution. Warning Signal.

Unacceptable performance. Action Signal.

*Pamela Christudoss*

Dr. Pamela Christudoss

CMC EQAS Coordinator

Christian Medical College, Vellore

Details:  
cmcre.ac.in  
Phone: +91-420-2283102

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not subcontract any components

\*\*\*\*\* End of Report \*\*\*\*\*

*[Handwritten Signature]*  
01/09/23

ASSISTANT SURGEON  
Govt. Primary Health Centre  
Sembakkam-603108.



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - JULY 2023



PC-1024

Lab Name: Govt. PHC, Sembakkam

Lab No: 10532

Constituent Group: Chemistry III

Date of Result Entered: 20/07/2023

PT Item: Lyophilized human serum based

Date of Report Published: 07/08/2023

S No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1876	109.35	5.55	0.07	103 mg/dL	-1.05	0.28
2	CREATININE III	Jaffers Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1850	5.83	6.00	0.35	3.0 mg/dL	-0.06	0.02
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1776	113.00	4.12	4.00	113 mg/dL	0.00	0.22
<b>SDI Range</b>										
				<b>Interpretation</b>						
Within -1.00 to +1.00				Excellent.						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

**LAB ADDRESS :**

Govt. PHC, Sembakkam  
CLINICAL LABORATORY KANCHIPURAM(DIST)  
Sembakkam  
TAMILNADU603108

Coordinator Contact Details:  
Email:kalinge@cmcvellore.ac.in  
Contact Number: 0416-2283102

*Pamela Christodoss*

Dr. Pamela Christodoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
Data in CMC EQAS reports is confidential  
CMC EQAS does not sub contract any components  
\*\*\*\*\* End of Report \*\*\*\*\*

*Dr. [Signature]*  
20/8/23  
ASSISTANT SURGEON  
Govt. Primary Health Centre  
Sembakkam

CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - JUNE 2023



Sembakkam Lab No 10532  
 III Date of Result Entered : 20/06/2023  
 d human serum based Date of Report Published : 04/07/2023

Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	L
				CV	SD			
POD III	Robonik ( Semi automatic / Fully Automatic )	1823	286.63	5.17	14.82	309 mg/dL	1.51	0.1
Kinetic - Alkaline Picrate	Robonik ( Semi automatic / Fully Automatic )	1848	2.57	6.64	0.17	2.6 mg/dL	0.18	0.1
-PAP III	Robonik ( Semi automatic / Fully Automatic )	1804	113.73	4.46	5.07	111 mg/dL	-0.54	0.1

Interpretation
Excellent.
Good.
Accept with caution. Warning Signal.
Unacceptable performance. Action Signal.

JRAM(DIST)

Contact Details:  
 Email: [qc@cmcvellore.ac.in](mailto:qc@cmcvellore.ac.in)  
 Number: 0416-2283102

*Pamela Christudoss*  
 Dr. Pamela Christudoss  
 CMC EQAS Coordinator  
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
 Data in CMC EQAS reports is confidential  
 CMC EQAS does not sub contract any components  
 \*\*\*\*\* End of Report \*\*\*\*\*

*[Signature]*  
**ASSISTANT SURGEON**  
 Govt. Primary Health Centre  
 Sembakkam-603108.



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - APRIL 2023



PC-1024

Lab Name: Govt. PHC, Sembakkam  
Constituent Group: Chemistry III  
PT item: Lyophilized human serum based

Lab No: 10232  
Date of Result Entered: 19/04/2023  
Date of Report Published: 00/03/2023

Sl. No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonic ( Semi automatic / Fully Automatic )	1832	119.43	6.04	7.21	120 mg/dL	0.91	0.34
2	CREATININE III	Jeffes Kinetic - Alkaline Phosphate III	Robonic ( Semi automatic / Fully Automatic )	1820	2.10	8.09	0.18	2.35 mg/dL	1.09	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonic ( Semi automatic / Fully Automatic )	1730	111.01	5.19	5.79	100 mg/dL	-0.97	0.28
<b>SDI Range</b>										
				<b>Interpretation</b>						
Within -1.00 to +1.00				Excellent.						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

LAB ADDRESS :  
Govt. PHC, Sembakkam  
CLINICAL LABORATORY, KANCHIPURAM(DIST)  
Sembakkam  
TAMILNADU603103

Coordinator Contact Details:  
Email:clinqc@cmvce.ore.ac.in  
Contact Number: 0410-2283102

*Pareela Christudass*  
Dr. Pareela Christudass  
CMC EQAS Coordinator  
Christian Medical College, Vellore

Homogeneity and stability of the sample is passed.  
Data in CMC EQAS reports is confidential  
CMC EQAS does not sub contract any components  
\*\*\*\*\* End of Report \*\*\*\*\*

*[Signature]*  
ASSISTANT SURGEON  
Govt. Primary Health Centre  
Sambakkam



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - MARCH 2023



Lab Name

Govt. PHC, Sembakkam

Lab No

10532

Constituent Group

Chemistry III

Date of Result Entered

18/03/2023

PT Item

Lipohyalized human serum based

Date of Report Published

04/04/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Year Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1878	91.48	7.46	0.82	52.3 mg/dL	0.12	0.31
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1851	1.33	9.08	0.12	1.28 mg/dL	-0.41	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1761	110.93	5.23	2.80	100.6 mg/dL	-0.75	0.28
<b>SDI Range</b>										
				<b>Interpretation</b>						
Within - 1.00 to +1.00				Excellent.						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.59				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

### LAB ADDRESS :

Govt. PHC, Sembakkam  
CLINICAL LABORATORY KANCHIPURAM(DIST)  
Sembakkam  
TAMILNADU603108

### Coordinator Contact Details:

Email:cmceqa@cmcvellore.ac.in  
Contact Number: 0410-2283102

*Pamela Christudoss*

Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
Data in CMC EQAS reports is confidential  
CMC EQAS does not sub contract any components

\*\*\*\*\* End of Report \*\*\*\*\*

**ASSISTANT SURGEON**  
Govt. Primary Health Centre  
Sembakkam-603108.