

NEW LAB MONTHLY SUMMARY

Lab Name: Govt. Block PHC, Acharapakkam  
 Lab No: T072  
 Month: March  
 Year: 2023  
 Constituent Group: Chemistry III

Details About Robust Analysis  
 Detail About Monthly Summary  
 Detail about SDI

Click on the analyte to view Graphical Data

[All Analyser Result](#) [Print](#)

Date of Result Entered: 11/03/2023

Date of Report Published: 04/04/2023

Sl No.	Analyte	Method - Principle Name	Analyser Name	No. of Measurements	CV	Participant			SDI	U
						En	SD	Value		
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1878	01.48	7.46	6.82	92 mg/dL	0.08	0.31
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1851	1.33	9.08	0.12	1.2 mg/dL	-1.07	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1761	110.93	5.23	5.80	114 mg/dL	0.53	0.28

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

<a href="#">Modify Address</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">View Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Logout</a>

*Par*

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. Block PHC, Acharapakkam  
 Lab No 7072  
 Month May  
 Year 2023  
 Constituent Group Chemistry III

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#)

[Print](#)

Date of Result Entered : 21/05/2023

Date of Report Published : 05/06/2023

Sl. No.	Parameter	Method	Instrument	Lab Result	Target	Delta	Delta 2	Delta 3	Delta 4	Delta 5	Delta 6
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1834	180.88	5.19	9.39	18.1	0.23	0.44	
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1854	1.76	7.48	0.13	1.7	-0.45	0.01	
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1798	114.60	5.08	5.83	101	-2.33	0.27	

SDI Range	Interpretation
Within -1.00 to -1.00	Excellent
Within ±1.01 to ±2.00	Good
Within ±2.01 to ±2.99	Accept with caution - Warning Signal
Beyond ±3.0	Unacceptable performance - Action Signal.

Links


**MEDICAL OFFICER**  
 UPGRATED GOVT. HEALTH CENTRE  
 ACHARAPAKKAM, PIN- 601 301  
 CHENGALPATTUR DISTRICT

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. Block PHC, Acharapakkam  
 Lab No 7072  
 Month April  
 Year 2023  
 Constituent Group Chemistry III

Details About Robust Analysis

Detail About Monthly Summary

Detail about SDI

Click on the analyte to view Graphical Data

All Analyser Result

Print

Date of Result Entered : 19/04/2023

Date of Report Published : 06/05/2023

1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1832	119.45	6.04	7.21	125 mg/dL	0.77	0.34
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1820	2.16	8.09	0.18	2.2 mg/dL	0.23	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1750	111.61	5.19	5.79	94 mg/dL	-3.04	0.28

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links


*H.V. S*  
**MEDICAL OFFICER**  
 UPGRADED GOVT. HEALTH CENTRE  
 ACHARAPAKKAM - 601 301  
 CHENNAI METRO DISTRICT

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. Block PHC, Acharapakkam  
 Lab No 7072  
 Month June  
 Year 2023  
 Constituent Group Chemistry III

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#)

[Print](#)

Date of Result Entered : 07/06/2023

Date of Report Published : 04/07/2023

Sl. No.	Analyte	Method	Instrument	Result	Target	Delta	Delta Z	Unit	SDI	Delta Z
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1823	286.63	5.17	14.82	292 mg/dL	-1.66	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1848	2.57	6.64	0.17	2.7 mg/dL	0.76	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1804	113.73	4.46	5.07	105 mg/dL	-1.72	0.24

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

Sl. No.	Link	Description
1	<a href="#">View Result</a>	View Result
2	<a href="#">View Summary</a>	View Summary
3	<a href="#">View Details</a>	View Details

*M. S. S. S.*  
**MEDICAL OFFICER**  
 UPGRADED GOVT. PRIMARY HEALTH CENTRE  
 ACHARAPAKKAM 603 301  
 CHENGALPATTU DISTRICT

[https://home.emcvellore.ac.in/diagz/ViewMonthlySummary/viewResult2\\_new.aspx?resDetId=266&noSamples=1&month=6&year=2023&groupId=9&y](https://home.emcvellore.ac.in/diagz/ViewMonthlySummary/viewResult2_new.aspx?resDetId=266&noSamples=1&month=6&year=2023&groupId=9&y)

*noSamples=1&month=6&year=2023&*

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. Block PHC, Acharapakkam  
 Lab No 7072  
 Month July  
 Year 2023  
 Constituent Group Chemistry III

Click on the analyte to view Graphical Data

Details About Robust Analysis  
 Detail About Monthly Summary  
 Detail about SDI

All Analyser Result

Date of Result Entered : 07/07/2023  
 Date of Report Published : 07/08/2023

Sl. No.	Analyte	Method	Instrument	Count	Mean	SD	CV	Target	SDI	CV
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1876	109.35	5.55	6.07	93 mg/dL	-2.69	0.28
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1850	5.83	6.00	0.35	5.8 mg/dL	-0.09	0.02
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1776	113.00	4.12	4.66	120 mg/dL	1.50	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

MEDICAL OFFICER  
 UPG-DEPT. GOVT. PHC HEALTH CENTRE  
 ACHARAPAKKAM - 603 301  
 CHENGALPATTUR DISTRICT

[https://home.cmcvellore.ac.in/cdnq/ViewMonthlySummary/viewResult2\\_new.aspx?resDetId=267&noSamples=1&month=7&year=2023&groupId=9&y](https://home.cmcvellore.ac.in/cdnq/ViewMonthlySummary/viewResult2_new.aspx?resDetId=267&noSamples=1&month=7&year=2023&groupId=9&y)

resDetId=268&noSamples=1&month=8&year=2023&groupId=9&y  
 - DISTRICT

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. Block PHC, Acharapakkam  
 Lab No 7672  
 Month August  
 Year 2023  
 Constituent Group Chemistry III

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#)

[Print](#)

Date of Result Entered : 19/08/2023

Date of Report Published : 01/09/2023

Sl. No	Analyte	Method	Instrument	Q1	Q2	Q3	Q4	Q5	Q6	Q7	Q8	Q9	Q10
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1841	175.58	5.13	9.00	183 mg/dL	0.82	0.42			
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1852	1.76	6.77	0.12	1.8 mg/dL	0.34	0.01			
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1764	113.80	4.07	4.64	109 mg/dL	-1.04	0.22			

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

---



---



---

*M.N. Of*  
 MEDICAL OFFICER  
 UPDHEALTH & FAMILY WELFARE CENTRE  
 ALHAKKATTA - 601301  
 CHENGALPATTU DISTRICT

[https://home.priwellcare.ac.in/clin/qc/ViewMonthlySummary/viewResult2\\_new.aspx?resDetId=268&noSamples=1&month=8&year=2023&groupid=9&y](https://home.priwellcare.ac.in/clin/qc/ViewMonthlySummary/viewResult2_new.aspx?resDetId=268&noSamples=1&month=8&year=2023&groupid=9&y)