



CALIBRATION REPORT

Report No : HBMS/CAL/23-24		Calibration Date : 21.09.2023
		Calibration valid: 20.03.2024
1.1.CUSTOMER DETAILS:		
Name and Address of the organization		M/s, PRIMARY HEALTH CENTRE, POOCHIYUR, COIMBATORE-641109.
1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):		
EQUIPMENT NAME		MICRO PIPETTE
i	Manufactured by	MICROLUX
ii	Model	
iii	Value	100 μ l-1000 μ l
iv	Serial No	PHC/MP/PO/001


Calibration standards used :

Master instrument	Model	Serial no	Date of calibration	Calibration due on
Electrical compact scale	SF-400C	SF400-2240	07-OCT-2022	06-OCT-2023

Calibration test Report

Number of measurement	Channel 1 (Volume)		
	100 μ l	500 μ l	1000 μ l
1	100.03	500.06	1000.14
2	100.04	500.09	1000.12
3	100.05	500.02	1000.10
4	100.09	500.08	1000.08
5	100.00	499.94	1000.09
6	100.08	500.07	1000.12

For Heart Beat Medical Systems,


(Authorized signatory)

