

**CALIBRATION REPORT**

Report No: HBMS/CAL/23-24		Calibration Date: 21.09.2023
		Calibration Valid: 20.03.2024
<b>1.1. CUSTOMER DETAILS:</b>		
Name and Address of the organization		<b>M/s, PRIMARY HEALTH CENTRE POOCHIYUR, COIMBATORE-641109</b>
<b>1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):</b>		
EQUIPMENT NAME		PIPETTE
I	Manufactured by	MICROLUX
II	Model	5µl - 50µl
III	Serial No	PHC/MP/PO/001
IV	Location	LABORATORY


**Calibration standards used :**

Master instrument	Model	Serial no	Date of calibration	Calibration due on
Electrical compact scale	SF-400C	SF400-2240	07-OCT-2022	06-OCT-2023

**Details of the machine:** machine calibrated and checked output.

Number of measurement	Channel 1 (Volume)		
	05µl	20µl	50µl
1	05.01	20.01	50.23
2	05.01	20.01	50.02
3	05.01	20.05	50.04
4	05.02	20.04	50.02
5	05.01	20.08	50.09
6	05.02	20.12	50.04

For Heart Beat Medical Systems,

  
(Authorized signatory)

