

			CALI	BRA	TION	CERT	<u> </u>	<u>E</u>		
CERTIFICATE NO	RTIFICATE NO: SBS/CL/23/		3/12295		MECHANICAL		ı	Page No:1 of 1		
Issue Date					30-09-20	023		•		
SRF No & Date				SRF/23/00501-0002 & 29-09-2023						
Receipt Date				29-09-2023						
Calibration Date					29-09-2023					
Calibration Due					28-09-2024					
Customer Name	& Addr	ess								
GOVERNMENT P	RIMAR	Y HEALTH	CENTRE,							
CHELLAMUTHU I	NAGAR	ZAMIN UTI	HUKULI,PO	LLACHI-6	42104.					
			De	tails of D	evice Un	der Calibra	ation (DUC)			
Description	escription : CENTRIFUGE			Make & Model		:	LABOTECH & N	IA		
Range	:	3500 RPM			Sr. No		:	76427		
Resolution	:	1	RPM		Identifica	ation No	;	NA		
DUC Condition	:	Satisfactor	у		Location			LABORATORY		
		Envi	ronmental	Condition	ıs & Star	dard Oper	ating Proced	ure Details		
Environmental Details Temperature: 2			re: 24.3 °C	24.3 °C Relative Hu		midity 54 % Rh				
Calibration Procedure No SBS/CP/ML/04			_/04	04 Calibration		done at ONSITE				
				Refer	ence Sta	ndards De	tails			
S.No Description	No Description Make/ S				SI No:		Certificate No		Validity	
1 Digital Tach	Digital Tachometer Lutron 8				& DT-2238		JRPM-CCTR-A&S-2022-0036		03-10-2023	
CALIBRATION RE	SULT	<u>s</u>	•	<u> </u>					-	

CALIDDATION OFFICIOATE

S.No	DEVICE UNDER CALIBRATION READINGS	STANDARD INSTRUMENTS	DEVIATION	EXPANDED UNCERTAINTY (±)	
	RPM	RPM	RPM	%	
1	1000	999.8	0.2	4.2	
2	2000	1999	1	4.2	
3	3000	2999	1	4.2	

REMARKS

- 1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2. The user should determine the suitability of the instrument for its intended use.
- 3. The recalibration interval should be determined on the user requirement.
- 4. The results stated in this certificate relate only to the item calibrated.
- 5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
- 6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

(Calibration Engineer) M.RAGUL





Chief Executive

Authorised Signatory

SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087, E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.