

			CALI	BRAT	ION	CERTI	FICAT	E		
CERTI	ERTIFICATE NO: SBS/CL/23/12388					MEG	CHANICAL	Pa	ge No:1 of 1	
ssue [Date	1		3	30-09-20	23				
SRF No & Date					SRF/23/00472-0007 & 30-09-2023					
Receipt Date					30-09-2023					
Calibration Date					30-09-2023					
Calibration Due					29-09-2024					
Custor	mer Name & A	ddress								
GOVE	RNMENT URB	AN PRIMARY H	EALTH CEN	NTRE,						
SADRA	AS,CHENGALF	ATTU DISTRIC	ЭΤ.							
			De	tails of Dev	vice Un	der Calibrati	on (DUC)			
Description : CENTRIFUGE			N	Make & Model :		LABORATORY CENTRIFUGE & NA				
Range : 6000 RPM			5	Sr. No			NA			
Resolution : 10 RPM				1	Identification No :			NA		
DUC Condition : Satisfactory				L	Location		:	LABORATORY		
		Envi	ronmental	Conditions	s & Star	idard Operat	ing Proced	ure Details		
Environmental Details Temperature: 2					24.3 °C Relative Humidity		54 % Rh			
Calibration Procedure No SBS/CP/ML/0				_/04	04 Calibration		done at ONSITE			
				Refere	nce Sta	ndards Deta	ils			
S.No	Description			Make/ SI	Make/ SI No:		Certificate No		Validity	
1	Digital Tachometer			Lutron &	Lutron & Q638380 JRPM-CCTR-			R-A&S-2022-0036	03-10-2023	
CALIB	RATION RESU	JLTS								
S.No	DEVICE UNDER CALIBRATION READINGS			DINSTRUM	INSTRUMENTS		ATION	EXPANDED UNCERTAINTY (±)		
RPM RPM			RPM	RPM		PM	%			
1	3000 2			2998.4	998.4		1.6	6		
2	5000			4998		2		6		
3	3 6000			5999		1			6	

1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.

2. The user should determine the suitability of the instrument for its intended use.

3. The recalibration interval should be determined on the user requirement.

4. The results stated in this certificate relate only to the item calibrated.

5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.

6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

(Calibration Engineer) C.SIVABALAN

Quality Manager (C.SIVABALAN)

Authorised Signatory

Chief Executive

SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087, E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.