

<b>—</b>												
				CAL	IBRAT	101	I CERT	<b>TIFIC</b>	47	<u>E</u>		
CER	<b>FIFICATE N</b>	O: SB	S/CL/23/	12386			M	EDICAL	DEVI	CES	Page	No:1 of 1
Issue	Date					30-09	9-2023					
SRF	No & Date					SRF/	23/00472-00	05 & 30-0	9-20	23		
Rece	ipt Date					30-09	9-2023					
Calibration Date						30-09-2023						
Calib	ration Due					29-09	9-2024					
Cust	omer Name	& Ad	dress									
GOV	ERNMENT	URBAI	N PRIMA	RY HEALTH	CENTRE,							
SADF	RAS,CHENC	GALPA	TTU DIS	TRICT.								
					Details of De	vice U	nder Calibra	tion (DUC	)			
Descr	iption	:	SEMI A		ER	Make	& Model	;		ROBONIK & PRIETES	ST TOUC	н
Range		:	MULTI		Sr. No		)	:		NA		
Resolution		:	MULTI			Identification No		:		NA		
DUC Condition : SA		SATIS	ACTORY		Location				LABORATORY			
				Environ	mental Condi	tions	& Calibration	Procedu	e De	tails		
Environmental Details Temperature:25.6 ° C					25.6 ° C	Relative Hun		nidity		52% RH		
Calibration Procedure No SBS				SBS/CP/MD/2	BS/CP/MD/20		Calibration d	one at		ONSITE		
					Refere	ence S	tandards Det	ails				
S.No	S.No Description Make/ SI No:					:		Certificate No			Validity	
1	Electrical Sa	afety Ar	nalyser		Rigel Medica	& 44L	-1059	M-23080	9-16-	4		10-08-2024
					•							L

## ELECTRICAL SAFETY

S.no	SPECIFICATION	MEASURED VALUES	EXPANDED UNCERTAINTY (±)		
1	Insulation Resisitance	Measured values in MΩ	Uncertainty in % ( ±)		
	>20MΩ	89	13.92		
2	Earth Leakage	Measured values in µA	Uncertainty in % (±)		
	<5000µAfor B,BF,CF	159	7.6		
3	Enclosure Leakage	Measured values in µA	Uncertainty in % (±)		
	<500µAfor B,BF,CF	226	7.2		

## REMARKS

RESULTS

1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.

2. The user should determine the suitability of the instrument for its intended use.

3. The recalibration interval should be determined on the user requirement.

4. The results stated in this certificate relate only to the item calibrated.

5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

(Calibration Engineer) C.SIVABALAN

Calibrated By

Technical Manager

Chief Executive

Authorised Signatory

## SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087, E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.