

			CAL	IBRAT	ION	I CERT	TIFIC	A	TE_	
CERTIFICATE NO: SBS/CL/23/12387				MEDICAL DEVICES					Page No:1 of 1	
Issue	Date				30-09	9-2023				
SRF	No & Date				SRF	23/00472-00	06 & 30	-09-	2023	
Rece	ipt Date				30-09	9-2023				
Calib	ration Date				30-09	9-2023				
Calib	ration Due	29-09-2024								
Cust	omer Name & A	Address								
GOV	ERNMENT URE	AN PRIMA	RY HEALTH	CENTRE,						
SADE	RAS,CHENGAL	PATTU DIS	STRICT.							
				Details of De	vice U	nder Calibra	tion (DL	C)		
Descr	intion :	ELEC	TRICAL SAFE	Y(CELL						
Desci	iption .	COUN	TER)		Make	& Model		:	SYSMEX & XP100	
Range	e :	MULT			Sr. N)			B0537	
Resol	ution :	MULT			Identi	fication No		:	NA	
DUC (Condition :	SATIS	FACTORY		Locat	ion		:	LABORATORY	
			Environi	mental Condi	tions	& Calibration	Proced	ure [Details	All a second
Enviro	onmental Details		Temperature:	25.6 ° C		Relative Hun	nidity		52% RH	
Calibr	ation Procedure I	No	SBS/CP/MD/2	29		Calibration d	one at		ONSITE	
				Refere	nce S	tandards Det	tails		•	
S.No	Description			Make/ SI No:	:		Certific	ate	No	Validity
1	Electrical Safety Analyser Rigel Med		Rigel Medica	cal & 44L-1059		M-230809-16-4			10-08-2024	

ELECTRICAL SAFETY

RESULTS

S.no	SPECIFICATION	MEASURED VALUES	EXPANDED UNCERTAINTY (±)
1	Insulation Resisitance	Measured values in MΩ	Uncertainty in % (±)
	>20MΩ	90	13.92
2	Earth Leakage	Measured values in µA	Uncertainty in % (±)
	<5000µAfor B,BF,CF	198	7.4
3	Enclosure Leakage	Measured values in µA	Uncertainty in % (±)
	<500µAfor B,BF,CF	234	7.2

REMARKS

- 1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2. The user should determine the suitability of the instrument for its intended use.
- 3. The recalibration interval should be determined on the user requirement.
- 4. The results stated in this certificate relate only to the item calibrated.
- 5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00 .
- 6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

Authorised Signatory

C.SIVABALAN

Technical Manager
C.SHANMUGARAJ

Chief Executive