

VIEW LAB MONTHLY SUMMARY

Details About Robust Analysis

Detail About Monthly Summary

Detail about SDI

Lab Name Govt. PHC, Thandampalayam  
 Lab No 12213  
 Month August  
 Year 2023  
 Constituent Group Chemistry III

Click on the analyte to view Graphical Data

All Analyser Result

Print

Date of Result Entered : 17/08/2023

Date of Report Published : 01/09/2023

Sl. No	Analyte	Method / Principle Name	Analyzer Name	No. of Participants	DV	CV	SD	Yours Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1841	175.58	5.13	9.00	170.2 mg/dL	-0.60	0.42
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1852	1.76	6.77	0.12	1.7 mg/dL	-0.50	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1764	113.80	4.07	4.64	104 mg/dL	-2.11	0.22
<b>Interpretation</b>										
SDI Range				Excellent.						
Within -1.00 to +1.00				Good.						
Within ±1.01 to ±2.00				Accept with caution. Warning Signal.						
Within ±2.01 to ±2.99				Unacceptable performance. Action Signal.						
Beyond ±3.0										

*A.K. Suresh*

**மருத்துவ அலுவலர்**  
 அரசு ஆர்திசு கலாநிதர் நிலையம்  
 தாமிலாந் மாநாஸம் - 638 109.  
 ஈக்டாடு மாநா.லிம்.

Links