

VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. PHC, Irumbedu	Details About Robust Analysis
Lab No	12217	Detail About Monthly Summary
Month	March	Detail about SDI
Year	2023	
Constituent Group	Chemiatry III	

Click on the analyte to view Graphical Data

[All Analyser Result](#) [Print](#)

Date of Result Entered : 17/03/2023

Date of Report Published : 04/04/2023

SLNo	Analyte	Method / Principle Name	Analyser Name	No of Participants	DV	Participants			Your Value	SDI	U
						CV	SD				
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1678	91.48	7.46	6.82	90.81 mg/dL	-0.10	0.31	
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1651	1.33	9.08	0.12	1.37 mg/dL	0.33	0.01	
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1761	110.93	5.23	5.80	112.79 mg/dL	0.32	0.28	

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

P. Ash
THE MEDICAL OFFICER
GOVT. PRIMARY HEALTH CENTRE
IRUMBEDU

Links