

## VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Inumbedu  
 Lab No 12217  
 Month April  
 Year 2023  
 Constituent Group Chemistry III

Details About Robust Analysis

Detail About Monthly Summary

Detail about SDI

Click on the analyte to view Graphical Data



Date of Result Entered : 17/04/2023

Date of Report Published : 06/05/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants			Your Value	SDI	U
						CV	SD	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1832	119.45	6.04	7.21	123.52 mg/dL	0.56	0.34	
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1820	2.16	8.09	0.16	2.3 mg/dL	0.80	0.01	
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1750	111.61	5.19	5.79	113.04 mg/dL	0.25	0.28	

  

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

*PAH*  
 THE MEDICAL OFFICER  
 GOVT. PRIMARY HEALTH CENTRE  
 INUMBEDU

## Links

[https://home.cmcvellore.ac.in/clinqc/ViewMonthlySummary/ViewResult2\\_new.aspx?resDetId=264&noSamples=1&month=4&year=2023&groupId=9&yearGroup=110](https://home.cmcvellore.ac.in/clinqc/ViewMonthlySummary/ViewResult2_new.aspx?resDetId=264&noSamples=1&month=4&year=2023&groupId=9&yearGroup=110)