



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - FEBRUARY 2023



PC-1024

Lab Name Govt. Block PHC, Parandur

Lab No 7070

Constituent Group Chemistry III

Date of Result Entered : 18/02/2023

Patient Lyophilized human serum based

Date of Report Published : 08/03/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1868	155.11	6.56	10.18	168.5 mg/dL	1.31	0.47
2	UREA III	Urease UV / GLDH III	Robonik (Semi automatic / Fully Automatic)	175	30.39	12.93	3.93	38.5 mg/dL	2.06	0.59
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1872	3.40	9.00	0.31	3.5 mg/dL	0.33	0.01
4	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1788	112.79	5.23	5.90	125.2 mg/dL	2.10	0.28
SDI Range Interpretation										
Within -1.00 to +1.00				Excellent						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

LAB ADDRESS :

Govt. Block PHC, Parandur

"The Block Medical Officer Parandur, Parandur Post, Karapettai via, Parandur, Kancheepuram District
Parandur
TAMILNADU631552

TAMILNADU631552

MEDICAL OFFICER,
Govt. Primary Health Centre
PARANDUR - 631 552
Kanchipuram District

Pamela Christudoss

Coordinator Contact Details:
Email: clingqc@cmcvellore.ac.in

Dr. Pamela Christudoss
CMC EQAS Coordinator