



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - MARCH 2023



PC-1024

Lab Name Govt. Block PHC, Parandur

Lab No

7070

Constituent Group Chemistry III

Date of Result Entered :

20/03/2023

PT item Lyophilized human serum based

Date of Report Published :

04/04/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	CV	Participants	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1878	91.48	7.46	6.82	100.3	1.29	0.31	
2	UREA III	Urease UV / GLDH III	Robonik (Semi automatic / Fully Automatic)	175	55.20	11.70	6.46	50.5	-0.73	0.98	
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1851	1.33	9.08	0.12	1.22	-0.91	0.01	
4	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1761	110.93	5.23	5.80	127.8	2.91	0.28	
Interpretation											
SDI Range				Excellent							
Within -1.00 to +1.00				Good.							
Within ±1.01 to ±2.00				Accept with caution. Warning Signal.							
Within ±2.01 to ±2.99				Unacceptable performance. Action Signal.							
Beyond ±3.0											

LAB ADDRESS :

Govt. Block PHC, Parandur
"The Block Medical Officer Parandur, Parandur Post, Karapettai via, Parandur, Kancheepuram District
Parandur
TAMILNADU631552

MEDICAL OFFICER,
Govt. Primary Health Centre
PARANDUR - 631 552
Kanchipuram District

Pamela Christudoss

Coordinator Contact Details:
Email:clinqc@cmcvellore.ac.in

Dr. Pamela Christudoss
CMC EQAS Coordinator