



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - APRIL 2023



PC-1024

Lab Name Govt. Block PHC, Parandur

Lab No 7070

Constituent Group Chemistry III

Date of Result Entered : 20/04/2023

PT Item Lyophilized human serum based

Date of Report Published : 06/05/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1832	119.45	6.04	7.21	127.7 mg/dL	1.14	0.34
2	UREA III	Urease UV / GLDH III	Robonik (Semi automatic / Fully Automatic)	169	42.32	10.68	4.52	36.5 mg/dL	-1.29	0.70
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1820	2.16	8.09	0.18	2.32 mg/dL	0.91	0.01
4	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1750	111.61	5.19	5.79	104.2 mg/dL	-1.28	0.28
SDI Range				Interpretation						
Within -1.00 to +1.00				Excellent.						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

LAB ADDRESS :

Govt. Block PHC, Parandur
"The Block Medical Officer Parandur, Parandur Post, Karapettai via, Parandur, Kancheepuram District Parandur
TAMILNADU631552

MEDICAL OFFICER,
Govt. Primary Health Centre
PARANDUR - 631 552
Kanchipuram District

Parvula Christodoss

Coordinator Contact Details:
Email:clinqc@cmcvellore.ac.in

Dr. Pamela Christudoss
CMC EQAS Coordinator