



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - MAY 2023



PC-1024

Lab Name Govt. Block PHC, Akkur  
Constituent Group Chemistry III  
PT Item Lyophilized human serum based

Lab No 7528  
Date of Result Entered : 22/05/2023  
Date of Report Published : 05/06/2023

| Sl.No | Analyte        | Method / Principle Name               | Analyzer Name | No of Participants | DV     | Participants |      | Your Value | SDI   | U    |
|-------|----------------|---------------------------------------|---------------|--------------------|--------|--------------|------|------------|-------|------|
|       |                |                                       |               |                    |        | CV           | SD   |            |       |      |
| 1     | GLUCOSE III    | GOD-POD III                           | Biosystems    | 33                 | 179.81 | 5.48         | 9.85 | 179 mg/dL  | -0.08 | 3.43 |
| 2     | CREATININE III | Jaffes Kinetic - Alkaline Picrate III | Biosystems    | 38                 | 1.78   | 8.91         | 0.16 | 1.9 mg/dL  | 0.75  | 0.05 |
| 3     | CHOLESTROL III | CHOD-PAP III                          | Biosystems    | 36                 | 116.50 | 6.42         | 7.47 | 117 mg/dL  | 0.07  | 2.49 |

| SDI Range             | Interpretation                           |
|-----------------------|--|
| Within -1.00 to +1.00 | Excellent.                               |
| Within ±1.01 to ±2.00 | Good.                                    |
| Within ±2.01 to ±2.09 | Accept with caution. Warning Signal.     |
| Beyond ±3.0           | Unacceptable performance. Action Signal. |

LAB ADDRESS :  
Govt. Block PHC, Akkur  
THE BLOCK MEDICAL OFFICER, CLINICAL LABORATORY, THIRUVANNAMALAI DISTRICT.  
Akkur  
TAMILNADU 631701

Coordinator Contact Details:  
Email: clinqc@cmcvellore.ac.in  
Contact Number: 0416-2283102

*V. G. P. H. C.*  
SS / SURGEON  
J.G.PHC AKKU  
PIN - 631 701

*Pamela Christudoss*  
Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
Data in CMC EQAS reports is confidential  
CMC EQAS does not sub contract any components  
\*\*\*\*\* End of Report \*\*\*\*\*